

Fixation Techniques in ACL Reconstruction

**Essay
Submitted by
Mohamed Saleh Hamdy
(M.B.,B.Ch)**

**For partial fulfillment of the Master Degree
In
"Orthopedic Surgery"**

Under supervision of

**PROF. DR. \ ABDEL AZIZ M. AL SINGERGY
Professor of Orthopedic Surgery
Faculty of Medicine
Cairo-University**

**DR.\ ABDALLAH MOHAMED AHMED
Lecturer of Orthopedic Surgery
Faculty of Medicine
Cairo-University**

Faculty of Medicine
Cairo-University
2010



إِنْ أُرِيدُ إِلَّا الْإِصْلَاحَ مَا اسْتَطَعْتُ وَمَا تَوْفِيقِي
إِلَّا بِاللَّهِ عَلَيْهِ تَوَكَّلْتُ وَإِلَيْهِ أُنِيبُ

سورة هود الآية (٨٨)





Acknowledgement

First of all thanks to Allah whose help was my main support in achieving this work.

I am intensely appreciated to ***Prof. Dr. Abdel Aziz M. Al Singergy***, Professor of Orthopedic Surgery, Faculty of Medicine, Cairo University, for his fatherly attitude and great effort.

Words fail me to express how grateful I am to ***Dr. Abdallah Mohamed Ahmed***, Lecturer of Orthopedic Surgery, Faculty of Medicine, Cairo University, for spending his time, effort and scientific experience in assisting this work.



Mohamed Saleh Hamdy



DEDICATION

Dedicated To ... 

*My **Parents** for their great help,*

*My **Wife***

*And my **Son and Daughter***



ABSTRACT

Fixation options of the ACL graft are numerous such as screws used a post, interference fixation, cress-pin ACL and endobutton for soft tissue femoral graft fixation. Double spike plate, washerLoc and Bone Dowel, Whipstitch-post and Fastlok device for soft tissue tibial graft fixation. Interference screw and endobutton for BPTB fixation.

Press fit technique without hardware fixation is recently added to graft fixation options.

KEY WORDS

ACL techniques

ACL reconstruction

Fixation acl

CONTENTS

<i>Title</i>	<i>Page</i>
INTRODUCTION	1
THE AIM OF THE ESSAY	2
REVIEW OF LITERATURE	
▪ Structural And Mechanical Properties Of Anterior Cruciate Ligament	3
▪ Biomechanical Characteristics of Proper ACL Reconstruction	14
▪ Fixation Techniques of ACL reconstruction	2٤
▪ Comparison Between Different Methods of Fixation of ACL	108
SUMMARY AND CONCLUSION	120
REFERENCES	12٢
ARABIC SUMMARY	

LIST OF FIGURES

<i>Title</i>	<i>Page</i>	
Fig. (1):	Because of the anatomical shape of the insertion areas at the femoral and tibial insertion site, anterior fibers of the ACL are able to bend around the anterior edge of the intercondylar notch. Common reconstruction techniques do not replicate the foot type insertion at the tibia. A straight graft (dotted lines) such as a bone–patellar tendon–bone or hamstring graft may impinge at the notch in knee positions close to extension	4
Fig. (2):	(A) Full extension, the medial femoral condyle is cut off to inspect the femoral attachment of the ACL. (B) 90° of flexion. Arthroscopic surgery is normally performed with the knee in 90° of flexion. In clinical literature the notch roof is considered to be superior (anatomic description: anterior), the anterior rim is considered as shallow, the posterior part is considered as deep, and the posterior part is considered as inferior.	5
Fig. (3):	Toluidine blue stained histological sections of the femoral origin. (A) The femoral origin has the structure of a chondral apophyseal enthesis consisting of 4 layers. (B) Close up of the enthesis. The 4 layers can be described as ligament fibers, nonmineralized cartilage zone, mineralized cartilage zone, and subchondral bone plate.	7
Fig. (4):	Blood supply of the ACL. The distribution of the blood supply is not homogenous. A vascular areas can be found at the insertion sites and within the area of fibrocartilage-	9

	like cells.	
Fig. (5):	Load deformation curve	10
Fig. (6):	Stress strain curve	12
Fig. (7):	Attachment of the ACL on the femur. (A) Dotted line shows the attachment of the main fibers of the ACL. AFS: A parallel line of the axis of the femoral shaft. (8) When we drew a vertical line (VL) through the contact point (C) between the femoral condyle and the tibial plateau on a picture taken at 90° of flexion, this, line and the long axis of the ACL attachment (AX) was crossed at the point (PL) on the vertical line 5 to 8 mm anterior to the edge of the joint cartilage. The center of the attachment of the anteromedial bundle (the AM point) was located at the point 5 to 6 mm distal from the back of the femur in measurement using the onset guide	16
Fig. (8):	Navigation for the tibial tunnels in arthroscopic surgery. (A) The concept of the Wire-navigator, which is composed or a Navi-tip (a) and a Wire-sleeve (b). The Navi-tip consisted of tibial indicator (c) and femoral indicator (d). The axis of the Wire-sleeve passed through the lip of the tibial indicator. The direction and position of the Wire-sleeve was automatically decided, independent of those or the Navi-i ip. (B) The Tibial-indicator of the Navi-tip is placed at the center or the PL bundle footprint on the tibia. Then, keeping the tibial indicator at this point, we aim the Femoral-indicator at the center of the postrolateral bundle footprint on the femur. (C) Arthroscopic visual field of the Nan-lip.	18
Fig. (9):	Two tibial tunnel outlets are shown. They are created at	19

	the center of the normal attachment of the AM and PL bundles, respectively	
Fig. (10):	How to insert a guide wire for the PL bundle into the femur using the transtibial tunnel technique. (A) first, a surgeon taps a Kirschner wire into the center of the PL bundle attachment on the femur, and, then, drills it into the femur (observation through the medial portal). (B) A cannulated drill inserted using the wire as a guide (observation through the medial portal). (C) Two tunnel outlets created on the femur (observation through the medial portal). (D) The reconstructed 2 bundles (observation through the lateral portal). The posterolateral bundle was observed behind the AM bundle. An absorbable suture marker is attached to each graft to show the point of tip of the Endobutton. (E) Anteroposterior picture of the 2 bundles. Note the difference in the direction between the 2 grafts.	21
Fig. (11):	The 4.5mm drill bit is drilled through the cortex over the previously drilled, laser-marked, long guide pin after the socket has been drilled.	26
Fig. (12):	The prepared graft is passed through the fabric loop until each arm of the graft is the same length	27
Fig. (13):	The colored absorbable monofilament suture locks the graft to prevent sliding and also marks the entry point for the femoral tunnel	28
Fig. (14):	All four sutures –two #5 and two #2 sutures are passed through the eye of the long passing pin	28
Fig. (14b):	A,B, The long guide pin is passed into the knee and out through the soft tissue with the graft construct attached	29

Fig. (15):	The #2 suture are used to flatten the endobutton, seating it on the femoral cortex after the graft has been pulled into the socket with the #5 suture.	30
Fig. (16):	Strong retrograde tension fails to dislodge the graft when the knee is cycled once the endobutton has been properly seated	30
Fig. (17):	The Xtendobutton fits over the endobutton so that it is large enough to hold outside a 10mm femoral tunnel.	31
Fig. (18):	in this revision case, the Xtendobutton was used to anchor the graft on the outside of a 9 mm femoral tunnel	32
Fig. (19):	A and B, Two –thirds threaded femoral cancellous screw and washer used as fixation post for whipstitches from four strands	35
Fig. (20):	A and B, Sutures through the endobutton eyelets are tied around the femoral screw post. The graft is looped through them	36
Fig. (21):	The EZLoc is a cortical femoral fixation device consisting of a gold-colored lever arm (a) connected to a slotted body (b) by an axle (c). The lever arm provides slippage-resistant, stiff, and strong fixation similar to a cross-pin because it seats on cortical bone (circle). The lever arm prevents proximal and distal micromotion, and the body prevents anteroposterior and medial lateral micromotion. The slot in the body accepts both one- and two-strand soft tissue anterior cruciate ligament grafts	37
Fig. (22):	The EZLoc comes in a sterile package with a passing pin and suture. The dull end of a 16-inch-long, sharp-tip passing pin is secured in the proximal end of the slotted body by a suture tied under tension. The passing pin is	39

	used to orient the lever arm laterally and to pull the anterior cruciate ligament (ACL) graft into place. In the special situation when the EZLoc is used in the posterolateral femoral tunnel in a two-tunnel ACL reconstruction technique and the posterolateral femoral tunnel is not in the same axis as the tibial tunnel, the passing pin is removed from the EZLoc and the suture is used to pull the graft.	
Fig. (23):	The length of the lateral wall of the femoral tunnel is measured with a depth gauge inserted through the tibial and femoral tunnels. The tip of the depth gauge is hooked on the lateral cortex of the femur. The length of femoral tunnel in this subject is 45mm and is measured at the opening of the femoral tunnel into the notch.	40
Fig. (24):	Stratis ST femoral fixation graft block and transverse locking pin.	41
Fig. (25):	Stratis ST femoral fixation system with soft tissue graft	42
Fig. (26):	Prepared graft placed through lower eyelet of graft block	43
Fig. (27):	A, Graft block-graft construct attached to graft inserter with transverse guide. B, graft block-graft construct attached	43
Fig. (28):	Switch stick through distal end of graft block	44
Fig. (29):	Drill sleeve advanced over switching stick	45
Fig. (30):	Fixation pin loaded onto inserter	45
Fig. (31):	Fixation pin advanced through upper eyelet of graft block	46
Fig. (32):	Close-up view of implant graft block with fixation pin	47
Fig. (33):	The Arthrex TransFix implant is made of titanium, and the Bio-TransFix implant is made of poly-L-lactic acid (PLLA	48

Fig. (34):	The absolute neurovascular: safe zone" during cross-pin guidewire placement is from +20 degrees (0 degrees equals "parallel to floor" line) to -40 degrees (lowering the guide more posteriorly)	49
Fig. (35):	Prepare the tibialis graft with a running or krackow stitch, and size it to the nearest-millimeter diameter	50
Fig. (36):	Complete the tibial tunnel through a small anterior medial tibial incision using a posterior cruciate ligament referencing guide	50
Fig. (37):	Remove the Beath pin. Insert the matched transfix tunnel hook through the tibial tunnel, and position it in the femoral socket	51
Fig. (38):	Drill the 5-mm broach with a stop collar over the 3-mm guidewire	52
Fig. (39):	Pass and deliver the nitinol wire out of the tibial tunnel. Pass the selected graft in a retrograde fashion	52
Fig. (40):	Seat the implant at the appropriate depth (match the calibration line to the reamer) with gentle taps using a mallet only after smooth passage of the wire is confirmed.	53
Fig. (41):	The transverse drill guide assembly ensures that the cross-pin will bisect the femoral tunnel; this is confirmed arthroscopically	55
Fig. (42):	The forked femoral guide pin is used to capture the Flexwire in the femoral tunnel and deliver it at the distal tibia	56
Fig. (43):	The graft is then looped over the Flexwire and advanced into the joint using the suture tape, which is passed through the forked femoral guide pin and brought out through the skin proximally	57

Fig. (44):	The Stryker Biosteon cross-pin is inserted over the guide pin	58
Fig. (45):	Position the Depuy Mitek Rigidfix femoral guide. Note that the arm of the guide is positioned lateral to the knee	59
Fig. (46):	Introduce the guide, and then drill the two cannulated sleeves	60
Fig. (47):	Arthroscopic inspection of the location of the Rigidfix holes	61
Fig. (48):	Different interference screws. A, Early metallic interference screw, sharp threaded; B, Sysorb biodegradable interference screw, round threaded; C, round-headed cannulated interference (RCI) screw, metallic, round threaded; megafix biodegradable interference screws, sharpthreaded at the tip, D, and round threaded at the body E.	63
Fig. (49):	Dilation of the screw site using a small dilator (4mm), followed by screw insertion	64
Fig. (50):	Notching of the anterior edge of the femoral tunnel entry to facilitate screw insertion and anterior placement of the screw. A, Schematic picture. B-G, intraoperative view	65
Fig. (51):	Bone wedge technique. At the anterosuperior position of the tunnel entry site, a thin bone wedge is detached using a special chisel (A). The screw then is advanced between the tunnel wall and this bone wedge	67
Fig. (52):	Specially designed chisel for the wedge technique	67
Fig. (53):	Crosspin system: graft harness (purple) and cross-pin, both with their leading sutures	68
Fig. (54):	Cross-pin drill bit as it engages the positioning rod	69
Fig. (55):	Use of an arthroscope in the transverse tunnel to verify	70

	that the implant's lead suture passes through the harness eyelet	
Fig. (56):	A, Tap the implant's driver to insert the cross-pin into the transverse tunnel. B, insert the implant until it is flush with the (Courtesy of ConMed-Linvatec).	70
Fig. (57):	Final view of the femoral fixation. The cross-pin engages the graft harness. (Courtesy of ConMed-Linvatec)	71
Fig. (58):	The ends of the polyester tape (C) are threaded through the buckle (A). The buckle is then flipped, followed by the insertion of the two prongs of the staple (B) attached at the end of the impactor (D).	74
Fig. (59):	The staple is position perpendicular to the tibial surface with the aid of the impactor (A) while maintaining tension on the final impaction to demonstrate the relation of the tape relative to the device: anteroposterior view (B), lateral view (C).	74
Fig. (60):	The bone plug obtained from the tibia during tunnel preparation (A) is inserted back into the tibial tunnel	75
Fig. (61):	Tibial screws for "whipstitch –post" fixation. A, /smith & Nephew, 4.5 mm, no washer used; B, Linvatec, 6.5mm, washer mandatory, C: Arthrex 4.5mm, Washer optional	76
Fig. (62):	Whipstitches are placed to fold the tendon on itself	77
Fig. (63):	With every two throws, the suture is pulled very tightly to eliminate slack as the tendon tubularizes	78
Fig. (64):	The WasherLoc is a low-profile washer with four long peripheral spikes and multiple short central spikes. The long peripheral spikes engage cortical bone and contain the soft-tissue anterior cruciate ligament graft under the washer. The multiple shorter spikes purchase the soft	79

	graft	
Fig. (65):	An 8-mm diameter bone dowel harvester is driven over a tibial guidewire to the level of the subchondral bone. The harvester is rotated several times to break the bone dowel away from the subchondral bone. The typical length of the bone dowel is 20 to 30 mm	80
Fig. (66):	The counterbore is oriented parallel to the back wall of the tibial tunnel and is aimed toward the fibula. The purpose of the counterbore is to recess the WasherLoc inside the tibial tunnel to avoid irritating the overlying soft tissues.	81
Fig. (67):	The awl (A) is threaded into the drill sleeve (B), which is threaded into the WasherLoc. Half of the graft is placed on either side of the awl. The WasherLoc is oriented parallel to the back wall of the tibial tunnel and is aimed toward the fibula.	81
Fig. (68):	The cone-shaped dilator is inserted anterior to the soft tissue anterior cruciate ligament graft and is impacted 25 mm into the tibial tunnel. The cone-shaped space prevents the cylindrical-shaped bone dowel from being driven into the joint.	82
Fig. (69):	The plastic sleeve is placed over the sharp tip of the bone dowel harvester and inserted into the dilated opening. The bone dowel is compacted into the tibial tunnel and anterior to the soft tissue anterior cruciate ligament graft	83
Fig. (70):	Double spike plate for graft fixation under an optional tension. The standard-sized DSP is a 1.5 mm-thick, 18-mm-long, 11-mm-wide, small plate made of titanium alloy with three holes and two spikes of 5.5mm length on the	84