



DOSIMETRIC STUDY OF FIELD JUNCTION IN ADJACENT BEAMS USING ASYMMETRIC COLLIMATORS AND MULTI LEAF COLLIMATOR (MLC)

By

Ahmed Zaki Zaki Youssef
B.Sc. in Biophysics, Ain Shams - University

**A thesis submitted in conformity with the requirement
for M.Sc. in Science (Physics - Biophysics),
Faculty of Women for Arts, Science and Education.**

SUPERVISORS

Prof. Dr. Hosnia M. Abu-Zeid
Prof. of Nuclear Physics
Phys. Dept., Women Faculty for
Arts, Sci., and Education,
Ain -Shams University.

Prof. Dr. Hoda A. Ashry
Prof. of Radiation Physics
National Center for Radiation
Research and Technology,
Atomic Energy Authority.

Dr. Khaled Mohamed El Shahat

Ass. Prof. of Medical Radiation
Biophysics,
Faculty of Medicine,
Al - Azhar University

2015

Supervision Committee

Approval Sheet

**Dosimetric Study of Field Junction in Adjacent
beams using asymmetric collimators and multi
leaf collimator (MLC)**

**by
Ahmed Zaki Zaki Youssef**

**Submitted for partial fulfillment of M.Sc. Degree
in Science (Physics - Biophysics),
Faculty of women for Arts, Science and Education.**

Approved by:

signature

**Prof. Dr. Hosnia M. Abu-Zeid
Prof. of Nuclear Physics
Phys. Dept., Women Faculty for Arts, Sci.,
and Education, Ain -Shams University.**

**Prof. Dr. Hoda A. Ashry
Prof. of Radiation Physics
National Center for Radiation Research and
Technology,
Atomic Energy Authority.**

**Dr. Khaled Mohamed El Shahat
Ass. Prof. of Medical Radiation Biophysics,
Faculty of Medicine,
Al-Azhar University**

بِسْمِ اللَّهِ الرَّحْمَنِ الرَّحِيمِ

قُلْ لَوْ كَانَ الْبَحْرُ مِدَادًا لِكَلِمَاتِ رَبِّي لَنَفِدَ الْبَحْرُ قَبْلَ

أَنْ تَنْفَدَ كَلِمَاتُ رَبِّي وَلَوْ جِئْنَا بِمِثْلِهِ مَدَدًا ﴿١٠٩﴾

قُلْ إِنَّمَا أَنَا بَشَرٌ مِثْلُكُمْ يُوحَىٰ إِلَيَّ أَنَّمَا إِلَهُكُمُ

إِلَٰهٌ وَاحِدٌ فَمَنْ كَانَ يَرْجُوا لِقَاءَ رَبِّهِ فَلْيَعْمَلْ عَمَلًا

صَالِحًا وَلَا يُشْرِكْ بِعِبَادَةِ رَبِّهِ أَحَدًا ﴿١١٠﴾

[سورة الكهف]

Acknowledgements

I am kneeling obsequiousness to ALLAH thanking HIM for showing me the right way. Without God help, my efforts would have gone astray. It was through the grace of God that I was able to acquire this great accomplishment. Thanks also for a person I love him very much, the Prophet Mohammed {God's praise and peace upon}.

Great thanks to the Head of Physics Department, Women Faculty for Arts, Science, and Education, Ain -Shams University, for her kind help and continuous encouragement for me and all young scientists in our department.

*Also, I wish to express all my deepest and sincerest gratitude to **Prof. Dr. Hosnia M. Abu-Zeid, Prof. of Nuclear Physics, Physics Depart., Women Faculty for Arts, Science, and Education, Ain - Shams University.** I am deeply indebted to her for her enormous patience, guidance, and support throughout the work program. My inspiration came from her passion for research and her confidence in my abilities. Working under her leadership in such an exciting field was both educationally and professionally enriching. As a model researcher, I look up to her for inspiration in my professional career.*

*I wish to express my deepest sincerest thanks to **Prof. Dr. Hoda Abdel-Moniem Ashry, Prof. of Radiation Physics, National Center for Radiation Research and Technology, AEA.** for her capable*

supervision, guidance that helped enhancing my understanding of this field and who has given me much guidance necessary for this project, and done so with a great deal of patience and valuable insight. She has been the driving force of this project.

*I would also like to extend a special thanks to **Dr. Khaled Mohamed El Shahat, Ass. Prof. of Medical Radiation Biophysics, Physics Depart., Faculty of Medicine, Al - Azhar University**, who was more helpful, understanding, and enjoyable to work with throughout the course of the work, and for his capable supervision, fruitful guidance, encouragement, ideas, endless help and many illuminating discussions through the course of the investigation.*

I would also like to thank the Radiation Biophysics Laboratory team, Physics Department, Women Faculty, Ain Shams University for supporting this work.

Finally, but in no way the least important, I would like to thank my family, for their support, encouragement, and understanding throughout my M.Sc.'s research and for all the good and bad times we had together.



Dedicated

To

*My parents (Father and
Mother),*

My wife,

My brother,

My new born son,

My professors

And

My colleagues

Contents

List of Symbols & Abbreviations.....	1
List of Figures.....	3
List of Tables.....	7
List of Published Paper.....	8
English Abstract.....	9

Chapter 1:

Theoretical Aspect

1.1	Radiotherapy Techniques.....	13
1.1.1.	How Does Radiation Work?.....	13
1.1.2.	Conventional external beam radiotherapy.....	16
1.1.3.	Conformal radiotherapy requires treatment planning.....	17
1.1.3.1.	Virtual simulation, 3-dimensional conformal radiotherapy, and intensity-modulated radiotherapy.....	19
1.1.3.2.	Planning requires high and verified accuracy.....	19
1.1.4.	Accuracy requirements for treatment planning calculations.....	20
1.1.4.1.	Fundamental approach.....	20
1.2.	Basic Physics of Radiotherapy.....	22
1.2.1.	Introduction.....	22
1.2.2.	Interaction properties of x-rays and electrons.....	23
1.2.3.	Clinical consideration for photon beams.....	25
1.2.3.1.	External beam radiation therapy (EBRT).....	25
1.2.3.2.	Effect of radiation in cancer treatment.....	29
1.2.3.3.	Effect of multiple beams.....	31
1.2.4.	Isodose curves.....	33
1.2.4.1.	Measurement of Isodose curves.....	37
1.2.4.2.	Parameters of Isodose curves.....	40
1.2.5.	Typical radiotherapy course.....	45
1.2.5.1.	Planning margins.....	49
1.2.5.2.	Specification of target dose.....	53
1.2.5.3.	The ICRU Reference Point.....	54
1.2.6.	Stationary Photon Beams.....	54

1.2.7. Rotation therapy	55
1.2.8. Dose volume histograms.....	55
1.2.9. Intensity-modulated radiation therapy delivery.....	58
1.2.9.1.Multileaf Collimator as Intensity Modulator.....	60

Chapter : 2

Materials and Methods

2. Materials.....	62
2.1. Medical Linear Accelerator:	62
2.1.1. Head construction and collimator system design	63
2.1.2. The treatment head and its components	65
Flattening filter	
Ionization chambers for dose monitoring	
Fixed Collimators	
Multi leaves Collimator	
2.2. Dosimetric instrumentation.....	68
2.2.1. The unidoser dosimeter.....	68
2.2.2. Ionization chambers	69
i)- Semiflex Chamber (0.125 cm ³) Type 31010, 31013.....	69
ii)- Pin point Ionization Chamber.....	70
Pinpoint Chamber (0.015 cm ³).....	70
2.2.3. Water phantom.....	71
2.2.4. Computed tomography.....	72
2.3. Methods	73
2.3.1. Beams Alignment and Position:-.....	73
2.3.1.1. Check of SSD and FSD Scales.....	73
2.3.2. Treatment planning system.....	74
2.3.3. Virtual Simulation and Field Setup.....	76
2.3.4. Contouring.....	77
2.3.5. Dose Calculation.....	77
2.3.6. Plan Evaluation.....	78
2.3.7. Commissioning and quality assurance.....	79

Results & Discussions

3	Introduction.....	84
3.1.	Effect of the Overlap Between Two Photon Fields on Dose Profile.....	84
3.1.1.	Dose profile for one photon beam (no overlap).....	84
3.1.2.	Dose profile for two adjacent photon beams (overlap).....	85
	Determination of Gap Separation and Abutment Region for Homogenous Dose Distribution.....	87
3.2.	Effect of Different Gap Separations on Dose Distribution Between Adjacent Fields.....	89
3.2.1.	Gap separation of about 1.0 mm.....	89
3.2.2.	Gap separation of about 5.0 mm.....	90
3.4	Effect of Different Abutment Region Widths on Dose Distribution.....	91
3.5.	The case of Cranio-Spinal Technique	93
3.6.	Case of Breast fields and supra-clavicular field matching.....	98
3.7.	Conventional Methods.....	102
3.8.	Three dimensional methods.....	102
3.8.1.	Three Dimensional Usual Technique.....	105
3.8.2.	Three Dimensional (3D) Asymmetric Technique.....	105
3.9.	SSD match without gap.....	111
3.9.1.	SSD Match with Gap.....	112
3.10.	Three Dimensional Usual.....	113
3.10.1	Three Dimensional Asymmetric.....	113
3.10.2.	Three Dimensional Non-Coplanar.....	114

Chapter 4 :

Conclusions.....	117
-------------------------	------------

References	119
Arabic Abstract	

List of Symbols & Abbreviations

AP	Anterior posterior
BSF	Backscatter Factor
CAX	Central Axis
CPE	Charged Particle Equilibrium
CTV	Clinical Target Volume
DVH	Dose-Volume Histogram
EBRT	External Beam Radiotherapy
IBRT	Internal Beam Radiotherapy
FS	Field Size
GR	Gantry Rotation
GA	Gantry Angle
GTV	Gross Tumor Volume
ICRU	International Commission on Radiation Units
IL	Isodose Line
IM	Internal Margin
IMRT	Intensity Modulated Radiotherapy
ISF	Inverse Square Factor
ITP	Inverse Treatment Planning
ITV	Internal Target Volume
IRTV	Irradiated Target Volume
LINAC	Linear Accelerator
MLC	Multi leaf Collimator
MU	Monitor Unit
OAD	Off-Axis Distance

List of Symbols & Abbreviations

OAR	Organ at Risk
PDD	Percentage Depth Dose
PA	Posterior Anterior
PP	Parallel Plate chamber
PTV	Planning Tumor Volume
RF	Radio-Frequency
RP	Radiation Photon
SAD	Source-Axis Distance
SCD	Source-Collimator Distance
SDD	Source -Diaphragm Distance
SF	Scatter Factor
SSD	Source-Surface Distance
TD	Target Dose
TL	Thermo-luminescence
TLD	Thermo-luminescent Dosimetry
TPS	Treatment Planning System
TV	Target Volume
VOI	Volumes of Interest

List of Figures

Chapter (1)

Theoretical aspects

Page No.

Fig. (1.1) : 3D treatment planning for a patient with a lung tumor.....	18
Fig. (1.2) : The dependence on photon energy and atomic weight for x-ray interaction with matter (khan 2010).....	25
Fig. (1.3) : Schematic diagram of a therapy x-ray tube with hooded anode, Khan, 2010.....	25
Fig. (1.4) : Multiple beams delivery, between each irradiation the gantry rotates to the next predefined angle.....	33
Fig. (1.5) : Isodose curve (Mould, 1981).....	34
Fig. (1.6) : Example of an isodose chart. A: Source to Surface Distance (SSD) type, Co-60 beam, SSD = 80 cm, field size = 10×10 cm at surface. B: Source to axis distance (SAD) type, ⁶⁰ Co beam, SAD =100cm, depth of isocenter = 10 cm, field size at isocenter=10×10 cm (Data from University of Minnesota Hospitals, Eldorado 8 Cobalt Unit, source size = 2 cm.), Khan, 2010.	35
Fig. (1.7) : Dose profile at depth showing variation of dose across the field. Co-60 beam, source - to - surface distance = 80 cm, depth 10 cm, field size at surface=10×10 cm. Dotted line indicates geometric field boundary at a 10cm depth.....	37
Fig. (1.8) : Isodose distributions for different - quality radiations. A : 200 kVp, source to surface distance (SSD)=50 cm, half -value layer = 1.0 mm Cu, field size=10×10 cm. B: ⁶⁰ Co, SSD=80 cm, field size=10x10 cm.....	41
Fig. (1.9) : Graphical representation of volume of interest as defined in “ICRU” reports No. 50 and 62,1999.....	52
Fig. (1.10): Dose volume histogram for ideal situation (the solid line) and for clinical situation (the dashed line).....	57

Chapter (2)

Materials and Methods

Page No.

Fig. (2-1) : DMX -Varian Medical Linear Accelerator.....	63
Fig. (2.2) : Collimator Assembly showed upper and lower collimator and optical field dimension / distance system.....	64
Fig. (2.3) : Movement of Collimator system.....	66
Fig. (2.4) : MLCs.....	67
Fig. (2.5) : Unidose Dosimeter	68
Fig. (2.6) : Semi flex Chambers Type 31010, 31013.....	69
Fig. (2.7) : Pinpoint Chamber (0.015 cm ³).....	70
Fig. (2. 8) : Water Phantom	71
Fig. (2.9) : CT Adaptive for Radiotherapy.....	73
Fig. (2.10): Inguinal nodes are targeted more precisely using fused CT/PET images.....	77
Fig. (2.11): Dose distribution of a proton beam to treat prostate cancer.....	78
Fig. (2.12): Surface dose rendering of a head and neck IMRT plan displays the dose distribution on the surface of the tumor, adjacent brain stem, and left parotid gland.....	79
Fig. (2.13): Eclipse displays percentage difference and distance-to-agreement values to compare	80

Chapter (3)

Results and Discussions

Fig. (3.1) : Relation between percentage relative dose vs. distance from fields junction for one photon beam individually. Photon has 6.0 MV energy and 20 cm x 20cm.....	86
• Fig. (3.2) : Relation between percentage relative dose vs. distance from fields junction for the case of two photon beams with 6.0 MV energy and 20 cm x 20 cm field size adjoined.....	87
• Fig. (3.3) : Relation between relative dose and the abutment regions widths for the matching techniques. Graphs show a change in structure between the 0.1 and 0.5 cm abutment region widths. Regions wider than 0.3 appear to have a similar curve structure.....	88

Fig. (3.4) : Relation between relative dose and abutment region width focused on the abutmen region widths ranged between 0.1cm to 1.0 cm.....	89
Fig. (3.5) : Measurements profiles of two adjacent photon beams, with 6.0 MV, and 20 cm x 20 cm field size, adjoined with 0.1 cm separation.....	90
Fig. (3.6) : Measurements profiles of two adjacent photon beams, with 6.0 MV and 20 cm x 20 cm field size, adjoined with 0.5 cm separation.....	91
Fig. (3.7) : Measurements profiles of two adjacent photon beams, with 6.0 MV and 20 cm x20 cm of field size, adjoined with 0.1 cm separation and then with 0.10 cm abu.....	92
Fig. (3.8) : Measurements Profiles of two adjacent photon beams with 6.0 MV and 20 x 20 cm field size, adjoined with 0.5 cm separation and then with 0.5 cm a.....	93
Fig. (3.9) : Treatment of patient position, volumes description and field arrangement.....	95
Fig. (3.10): Technique uses the same field arrangement as the classical Craneo-Spina (CS) technique (two lateral brain fields, 1–2 posterior spine fields)).....	95
Fig. (3.11): Skin gap between the spine fields rotated to AP orientations.....	96
Fig. (3.12): Calculation of collimator rotation.....	97
Fig (3.13): Dose profile for field junctions before applying collimator rotation and the gap.....	97
Fig. (3.14): Dose profile through junction after collimator rotation and ga.....	98
Fig. (3.15): Field matching between breast and supra-clavicular field in planning system.....	99
Fig. (3.16): Dosimetry evaluation for matching and hot-and cold-spots in matching gap (a) and dosimetric film for two-segment asymmetric breast field (b) Superposition of calculated (plain) and measured (dashed) iso- doses.....	100
Fig. (3.17): Relative dose in gap between breast and supra- clavicular.....	101
Fig. (3.18): Axial Plan for Right Breast cancer in TPS.....	102

Fig. (3.19): Axial , coronal and sagital views for gap between three fields technique for breast cance.....	104
Fig. (3.20): Plan sum for Breast and Supra-clav.....	105
Fig. (3.21): 3D display for matching between breast fields and supra-clav field.....	106
Fig. (3.22): Dose profile for the gap between supra-clav field and breast fields.....	107
Fig. (3.23): Dose profile for SSD matches without gap and 3D usual techniques.....	108
Fig. (3.24): dose profile for 3D asymmetric and 3D non-coplanar techniques.....	109
Fig. (3.25): Dose Volume Histogram (DVH) and Dose display for different organs.....	110
Fig. (3.26): dose profile at depth 2.0 cm without gap.....	112
Fig. (3.27): Dose profile at depth 2.0 cm with gap.....	112
Fig. (3.28): Dose profile for 3D usual techniques.....	113
Fig. (3.29): Dose profile for 3D asymmetric technique.....	114
Fig. (3.30): Dose profile for 3D non-coplanar technique.....	115