VALIDATION OF TWIN AXIAL ELECTRO-GONIOMETER IN ANGULAR MEASUREMENT

Thesis

Submitted in Partial Fulfillment for the Requirement of Master Degree in Physical Therapy

BY RANIA REDA MOHAMED ABDOU

B.Sc. in Physical Therapy Demonstrator of Physical Therapy for Basic science Department

SUPERVISORS

Prof. Dr. Mohsen Mohamed El-Sayyad
Professor of Physical Therapy for
Basic Science Department and
Director of Motion Analysis Lab.
Faculty of Physical Therapy
Cairo University

Ass. Prof. Dr. Amal Fawzy Ahmed
Assistant Professor of Physical Therapy
For Basic Science Department
Faculty of Physical Therapy
Cairo University

Dr. Neveen Abdel Latef Abdel Raoof


Lecturer of Physical Therapy

For Basic Science Department

For the of Physical Therapy

Faculty of Physical Therapy
Cairo University

Faculty of Physical Therapy
Cairo University
2009



ACKNOWLEDGEMENT

First of all, I would like to kneel thanking our God who provided me with essential power and patience for completion of this work.

I wish to express my deepest gratitude to **Prof. Or Mohsen El** Sayyad, Prof. of Physical Therapy, Director of Motion Analysis Lab., Basic Science Department, Faculty of Physical Therapy, Cairo University for kind supervision, generous support, unlimited helpful consultation, integration of ideas and continuous generous guidance through the preparation and conduction of this work.


I would like to appreciate special thanks for Assist. Prof. Dr.Amal Fawzy, Assist. Prof. of Physical Therapy, Basic Science Department, Faculty of Physical Therapy, Cairo University for kind supervision and encouragement to conduct this study.

I would like to appreciate special thanks for **Or.Neveen Abdel**Latef, lecturer of Physical Therapy Basic Science Department,

Faculty of Physical Therapy, Cairo University for kind supervision

and encouragement to conduct this study.

Special thanks to my professors of Basic science Department for their encouragement and to my colleagues in Faculty of Physical Therapy who helped me during the practical part of this work.



Validation of Twin Axial Electrogoniometer in angular measurement, Rania Reda Mohamed Abdou, Supervisors, Prof. Dr. Mohsen M. El Sayyad .Cairo University, Assist. Prof. Dr. Amal Fawzy .Cairo University, Dr. Neven Abdel Latef .Cairo University 2009 (Master Degree) 125 Pages.

Abstract

Background: Range of Motion measurement in Physical Therapy is very important to allow recording the state of the joint motion. Purpose: this study investigates the validity of a twin axial electrogoniometer in measuring joint angles (hip and knee). Methodology: Thirty normal subjects were assigned into one group, all the subjects were measured by the use of the electrogoniometer and the 3 Dimensional Motion Analysis system to measure Hip and Knee angles during sit to stand movement. Results: The study revealed that there was no statistical significant difference level in hip and knee joint angles between the electrogoniometer and the 3 Dimensional Motion Analysis systems during sit to stand movement. Discussion: The electrogoniometer may be considered as valid instrument in measuring joint Range of Motion during functional activity of daily living.

Key Words: Range of motion, Electrogoniometer, Three dimensional motion analyses, Sit to stand.






Table of Contents

Contents	Page
Chapter (I) Introduction	1
Statement of the problem	4
Purpose of the study	5
Justification of the study	6
Delimitation of the study	7
Limitation of the study	7
Hypothesis	8
Basic Assumption	8
Terminology	8
Chapter (II) Literature Review	11
Range of Motion	11
Measurement of Range of Motion	12
Biomechanics of sit to stand	35
Strategies affecting on sit to stand	39
Measurement of sit to stand	45
Summary of Literature Review	52
Chapter (III) Materials and Methods	54
Design of the study	54
Selection of the subject	54
Instrumentation	55
Testing Procedure	62
Data Collection	68
Data analysis and Statistical design	69
Chapter (IV) Results	70
Chapter (V) Discussion	85
Chapter (VI) Summary, Conclusion, Findings,	94
Implementations, and Recommendations	,
References	98
Appendices	120
Arabic Summary	



List of Tables

Table No	Page
(1) Comparison of Test book references for knee Active Range of Motion (in degree)	14
(2)Physical Characteristics of the subjects (Age –Weight – Height –BMI)	70
(3) Mean, Standard deviation and P value of the hip joint during flexion phase measured by electrogoniometer and 3DMA system	73
(4) Correlation Analysis between the angles of hip joint during flexion phase measured by electrogoniometer and 3DMA system	74
(5) Mean, Standard deviation and P value of the knee joint during flexion phase measured by electrogoniometer and 3DMA system	75
(6) Correlation Analysis between the angles of knee joint during flexion phase measured by electrogoniometer and 3DMA system	76
(7) Mean, Standard deviation and P value of the hip joint during extension phase measured by electrogoniometer and 3DMA system	77
(8) Correlation Analysis between the angles of hip joint during extension phase measured by electrogoniometer and 3DMA system	78.
(9) Mean, Standard deviation and P value of the knee joint during extension phase measured by electrogoniometer and 3DMAsystem	79
(10) Correlation Analysis between the angles of knee joint during extension phase measured by electrogoniometer and 3DMA system	80
(11) Mean, Standard deviation and P value of the hip joint during stand to sit phase measured by electrogoniometer and 3DMA system	18
(12) Correlation Analysis between the angles of hip joint during stand to sit phase measured by electrogoniometer and 3DMA system	82




(13) Mean ,Standard deviation and P value of the knee joint	
during stand to sit phase measured by electrogoniometer	83
and3DMAsystem	
(14) Correlation Analysis between the angles of knee joint	
during stand to sit phase measured by electrogoniometer	84
and3DMAsystem	
(15) Summary table of the hypothesis	93

.


٠

. -



List of Figures

Figure No	Page
(1) Phases of sit to stand	38
(2) The electrogoniometer used in the study	57
(3) The arms of the electrogoniometer	57
(4)The camera system and the chair used in the study	58
(5) Wand Kit used in the study	59
(6) The markers used in the study	60
(7)The modified chair used in the study	61
(8) The metronome used in the study	61
(9) The electrogoniometer placement	64
(10) The calibration set up procedure	65
(11)The markers placement	67
(12) Mean and Standard deviation of age of the subjects	71
(13) Mean and Standard deviation of Weight of the subjects	71
(14) Mean and Standard deviation of Height of the subjects	72
(15) Mean and Standard deviation of BMI of the subjects	72
(16) Mean and Standard deviation of the hip joint during flexion phase measured by electrogoniometer and 3DMA	73
(17) Correlation Analysis between the angles of hip joint during flexion phase measured by electrogoniometer and 3DMA	74
(18) Mean and Standard deviation of the knee joint during flexion phase measured by electrogoniometer and 3DMA	75
(19) Correlation Analysis between the angles of knee joint during flexion phase measured by electrogoniometer and 3DMA	76
(20) Mean and Standard deviation of the hip joint during extension phase measured by electrogoniometer and 3 DMA	77
(21) Correlation Analysis between the angles of hip joint during extension phase measured by electrogoniometer and 3DMA	78
(22) Mean and Standard deviation of the knee joint during extension phase measured by electrogoniometer and 3DMA	79



(23) Correlation Analysis between the angles of knee joint during extension phase measured by electrogoniometer and 3DMA	80
(24) Mean and Standard deviation of the hip joint during stand to sit phase measured by electrogoniometer and 3DMA	81
(25) Correlation Analysis between the angles of hip joint during stand to sit phase measured by electrogoniometer and 3DMA	82
(26) Mean and Standard deviation of the knee joint during stand to sit phase measured by electrogoniometer and 3DMA	83
(27) Correlation Analysis between the angles of knec joint during stand to sit phase measured by electrogoniometer and 3DMA	84