

# **Comparison of four novel customized implant abutments fabricated by pressing technology**

A thesis submitted to crown and bridge department, Faculty of Dentistry, Ain Shams University for the partial fulfillment of the doctoral degree requirements in Fixed Prosthodontics.

**BY**

**Marwa Mahmoud Emam**

B.D.S (Ain Shams University, 2007)

M.Sc (Ain Shams University, 2013)

Assistant Lecturer of Fixed Prosthodontics.

Crown and Bridge department.

Faculty of Dentistry, Ain Shams University.



## **Supervisors**

### **Dr/ Tarek Salah Morsi**

Assistant Professor and Head of Fixed Prosthodontics,  
Crown and Bridge department,  
Faculty of Dentistry, Ain Shams University.

### **Dr/ Ayman Galal El Dimeery**

Lecturer of Fixed Prosthodontics,  
Crown and Bridge department,  
Faculty of Dentistry, Ain Shams University.



## **Acknowledgement**

*I've always believed that writing the acknowledgement is harder than writing the whole thesis. There are so many great people that I want to thank and I am afraid there are no sufficient words to convey the level of gratitude I have for them.*

*My deepest gratitude is to my supervisor, **Ass. Prof. Dr. Tarek Salah El Din Morsi** for his guidance and encouragement over the years. He gave from his scientific knowledge, experience, effort and most of all his valuable time. I am grateful to his insightful comments and discussions that helped me develop and understand research protocol. I appreciate his unfailing attention to details. He was very tolerant and determined to see me through. He was such a wonderful motivation even when the road seemed tough for me. I will be always indebted to him. It was a great honor to work under his supervision.*

*I wish to express my sincere appreciation and gratitude to my supervisor, **Dr. Ayman Galal El-Dimeery** for years of mentorship, support and friendship. I have been amazingly fortunate to have him as a supervisor. He made me believe I had so much courage and strength to preserve even when I felt lost myself. I am grateful to him for holding me to high research standards. His careful readings and countless revisions are the reason of conducting this research despite all difficulties. He has always been there to answer any questions and clear any doubts. He has set an example of excellence as a researcher, mentor, instructor and role model. I aspire to emulate him.*

*Personal appreciation and thanks to **Dr. Amina Mohamed Hamdy, Dr. Amr Saleh EL-Etreby and Dr. Marwa Mohamed Wahsh** for their support, kindness and continuous encouragement over the years.*

*I wish to extend my sincere gratitude to my brother **Eng. Mostafa Mahmoud Emam** for his time and effort drawing the diagrams used in this thesis. Also I can't forget the talented **Dr. Karim tarek** and **Dr. Kareem Nagi** for offering their amazing photography skills whenever asked.*

*I am forever thankful to those who make the good times better and the hard times easier **Fatma Adel, Soha Osama, Sarah Mahanna, Hoda Sadek, Maha Salah, Heba Darwish, Menna Darwish, Noha Fayad, Nahla Hamed and Mai Mamdouh**. They are like the stars, that I know they will be always there to shine when it goes dark. They are the family I chose for myself.*

*I wish to extend my deep and sincere gratitude to my uncle **Eng. Osama Taha** for always being there when things crash down. Your love and encouragement from the early start have always been the motive behind my persistence.*

*I also extend my heartfelt gratitude to my students who made me find out that all that time trying to inspire them and they are the ones ended up inspiring me.*

*Last but by no means least, my deepest gratitude goes to my family for their unconditional love and support throughout my life and my studies. I*

*am forever indebted to them for giving me the opportunities and experiences that have made me who I am. This journey would not have been possible if not for them, and I dedicate this achievement to them.*

# **TABLE OF CONTENTS**

Introduction	1
Review of Literature	3
Statement of the Problem	33
Aim of the Study	34
Materials and Methods	35
Results	83
Discussion	107
Summary and Conclusion	120
References	124
Arabic Summary	



# LIST OF FIGURES

<b>Fig 1:</b> IPS e.max Press ingots LT .....	36
<b>Fig 2:</b> Sirona Ti Base, screw and scan body.....	36
<b>Fig 3:</b> Ultradent Porcelain etch (above), and Ultradent Silane (below). ....	37
<b>Fig 4:</b> TotalCem self etching self adhesive resin cement .....	37
<b>Fig 5:</b> Flowchart showing an overview of the study design CHAC: Cemented hybrid abutment crown, CHA: Cemented hybrid abutment, PHAC: Pressed hybrid abutment crown, PHA: Pressed hybrid abutment....	40
<b>Fig 6:</b> Schematic drawing of HAC of an upper central incisor showing structure composition and dimensions from proximal (left) and labial (right) view: 1: Ti Base, 2: e.max abutment crown as single restoration. ....	41
<b>Fig 7:</b> Schematic drawing of HA of an upper central incisor showing structure composition and dimensions from proximal (left) and labial (right) view: 1: Ti Base, 2: e.max hybrid abutment, 4: e.max allceramic crown. ....	41
<b>Fig 8:</b> Implant replica .....	42
<b>Fig 9:</b> Alignment of implant replica with the surveyor .....	43
<b>Fig 10:</b> Steady insertion of the abutment-replica combination into the doughy resin.....	44
<b>Fig 11:</b> Implant replica supported in tray resin .....	44
<b>Fig 12:</b> Schematic drawing of implant replica in the resin block .....	45
<b>Fig 13:</b> Stages of wax pattern buildup guided by the putty index .....	47
<b>Fig 14:</b> The putty indices for the waxed crown.....	47
<b>Fig 15:</b> Completed wax pattern on a Ti base screwed to the implant analog	48
<b>Fig 16:</b> The HAC wax pattern after margination with cervical wax .....	49
<b>Fig 17:</b> The sprued wax patterns of HACs (above) and HAs (below) attached to the ring base with the knotted floss inside their screw channels. ....	50
<b>Fig 18:</b> The investment ring before (left) and after setting and removing the silicone ring (right). ....	52

<b>Fig 19:</b> The investment ring inside the preheating furnace .....	52
<b>Fig 20:</b> IPS Alox Plunger and separator .....	54
<b>Fig 21:</b> Dropping the e.max ingots into the investment ring .....	54
<b>Fig 22:</b> The investment ring in the furnace .....	55
<b>Fig 23:</b> Marking the separation line with an Alox Plunger .....	55
<b>Fig 24:</b> The investment ring separated according to Alox plunger length ...	56
<b>Fig 25:</b> The partially blasted HAC samples .....	56
<b>Fig 26:</b> Completely divested (HA) samples (left) and (HAC) samples after divesting and separating the sprues. ....	57
<b>Fig 27:</b> PICO-MARK water-soluble pressure indicating paint. ....	58
<b>Fig 28:</b> Ti base painted with pressure indicating paint. ....	58
<b>Fig 29:</b> Optimum fit between the Ti base and the pressed restoration. ....	59
<b>Fig 30:</b> Construction of crowns wax pattern guided by the putty index .....	60
<b>Fig 31:</b> Crowns finished, marked and ready for pressing .....	60
<b>Fig 32:</b> Torque wrench kit .....	61
<b>Fig 33:</b> Mounted Ti bases before and after air blasting .....	62
<b>Fig 34:</b> Treating HAC with HF acid. ....	63
<b>Fig 35:</b> Treating HAC with silane coupling agent .....	63
<b>Fig 36:</b> Injection of resin cemeent inside a HAC .....	64
<b>Fig 37:</b> The sample in the loading device .....	65
<b>Fig 38:</b> The cemented samples on their corresponding Ti bases; HAC (above) and HA (below). ....	66
<b>Fig 39:</b> Treating the bonded surfaces of HA with HF .....	67
<b>Fig 40:</b> Treating the bonded surfaces of the crown with HF .....	67
<b>Fig 41:</b> Treating the bonded surfaces of the crown with silane coupling agent .....	68
<b>Fig 42:</b> Curing the cement at crown margins for 40 seconds. ....	69
<b>Fig 43:</b> Finished CHA sample with the crown bonded to the abutment. ....	69

<b>Fig 44:</b> Dissembled finished PHAC sample .....	70
<b>Fig 45:</b> Screwed finished PHAC sample .....	71
<b>Fig 46:</b> Finished PHA samples ready for investing.....	71
<b>Fig 47:</b> The PHACs attached to the ring base ready for investing (above) and after placing the silicon ring (below). ....	72
<b>Fig 48:</b> The P samples after pressing and divesting, HACs (above) and HAs (below). ....	73
<b>Fig 49:</b> The PHAC samples after separation .....	74
<b>Fig 50:</b> PHACs in the Invex liquid .....	75
<b>Fig 51:</b> Polishing of the Ti base to remove the oxide layer .....	75
<b>Fig 52:</b> Finished PHACs.....	76
<b>Fig 53:</b> Screwing the samples to the mounted analogs with the torque wrench .....	77
<b>Fig 54:</b> Schematic drawing of the HAC (left) and HA (right) either cemented or pressed ready for testing. ....	78
<b>Fig 55:</b> The custom holder used to hold the samples during static load test with the plastic cylinder inside (top view) .....	79
<b>Fig 56:</b> The custom holder used to hold the samples during static load test showing the 30 degrees inclination (lateral view) .....	80
<b>Fig 57:</b> The loading cell.....	80
<b>Fig 58:</b> Universal testing machine .....	81
<b>Fig 59:</b> Bar chart showing fracture resistance mean values for all groups separately (HAC: hybrid abutment crown, HA: hybrid abutment). ....	85
<b>Fig 60:</b> Line chart showing estimated marginal means of fracture load in relation to the pressing techniques with different designs (HAC: hybrid abutment crown, HA: hybrid abutment. ....	85
<b>Fig 61:</b> Bar chart showing fracture resistance mean values of different pressing techniques regardless restoration design. ....	87

<b>Fig 62:</b> Bar chart showing fracture resistance mean values of restoration design .....	88
<b>Fig 63:</b> Bar chart showing fracture resistance mean values of CHAC and PHAC (CHAC: Cemented hybrid abutment crowns, PHAC: Pressed hybrid abutment crowns). .....	91
<b>Fig 64:</b> Bar chart showing fracture resistance mean values of CHAC and CHA (CHAC: Cemented hybrid abutment crowns, CHA: Cemented hybrid abutment). .....	92
<b>Fig 65:</b> Bar chart showing fracture resistance mean values of CHAC and PHA (CHAC: Cemented hybrid abutment crowns, PHA: Pressed hybrid abutments). .....	93
<b>Fig 66:</b> Bar chart showing fracture resistance mean values of PHAC and PHA (PHAC: Pressed hybrid abutment crowns, PHA: Pressed hybrid abutment). .....	94
<b>Fig 67:</b> Bar chart showing fracture resistance mean values of PHAC and CHA (PHAC: Pressed hybrid abutment crowns, CHA: Cemented hybrid abutments). .....	95
<b>Fig 68:</b> Bar chart showing fracture resistance mean values of CHA and PHA (CHA: Cemented hybrid abutments, PHA: Pressed hybrid abutments). .....	96
<b>Fig 69:</b> Pie chart representing the mode of failure of cemented HAC group. ....	97
<b>Fig 70:</b> Pie chart representing the mode of failure of pressed HAC group. .	97
<b>Fig 71:</b> Pie chart representing the mode of failure of cemented HA group.	98
<b>Fig 72:</b> Pie chart representing the mode of failure of pressed HA group. ....	98
<b>Fig 73:</b> Pie chart representing the mode of failure of cemented group regardless the design. ....	99
<b>Fig 74:</b> Pie chart representing the mode of failure of pressed group regardless the design. ....	99

<b>Fig 75:</b> Stacked column chart showing mode of failure of cemented and pressed on groups.....	100
<b>Fig 76:</b> Pie chart representing the mode of failure of HAC group regardless the technique. ....	100
<b>Fig 77:</b> Pie chart representing the mode of failure of HA group regardless the technique. ....	101
<b>Fig 78:</b> Stacked column chart showing mode of failure of cemented and pressed group. ....	101
<b>Fig 79:</b> A fragment of a CHAC sample showing the resin cement layer attached to the surface.....	102
<b>Fig 80:</b> A fragment of a CHA sample showing the resin cement layer attached to the surface.....	103
<b>Fig 81:</b> The Ti base used in CHAC sample showing almost clean surface without resin cement attached to its surface. ....	103
<b>Fig 82:</b> A fragment of a PHAC sample showing the discolored surface in contact with the Ti base. ....	104
<b>Fig 83:</b> The Ti base used in PHAC sample showing the discolored black surface. ....	105
<b>Fig 84:</b> Sample showing fractured screw inside the Ti Base.....	106
<b>Fig 85:</b> Sample showing fractured screw inside the deformed implant replica at 50X.....	106

## LIST OF TABLES

<b>Table 1:</b> List of materials: .....	35
<b>Table 2:</b> Samples grouping .....	39
<b>Table 3:</b> Descriptive Statistics of fracture resistance of fracture resistance results measured in Newton (N) for different fabrication techniques and restoration designs.....	84
<b>Table 4:</b> Two-Way ANOVA for the effect of pressing technique and restoration design on the fracture resistance of custom hybrid abutment restorations.....	86
<b>Table 5:</b> Mean and standard deviation values of fracture resistance for different techniques regardless the design type. ....	87
<b>Table 6:</b> Mean and standard deviation values of fracture resistance for different designs regardless the pressing technique. ....	88
<b>Table 7:</b> One way ANOVA test. ....	89
<b>Table 8:</b> Results of the Tukey HSD Post Hoc test .....	90
<b>Table 9:</b> Results of the Tukey HSD Post Hoc test for fracture resistance mean values of CHAC and PHAC (CHAC: Cemented hybrid abutment crowns, PHAC: Pressed hybrid abutment crowns). ....	91
<b>Table 10:</b> Results of the Tukey HSD Post Hoc test for fracture resistance mean values of CHAC and CHA (CHAC: Cemented hybrid abutment crowns, CHA: Cemented hybrid abutment).....	92
<b>Table 11:</b> Results of the Tukey HSD Post Hoc test for fracture resistance mean values of CHAC and PHA (CHAC: Cemented hybrid abutment crowns, PHA: Pressed hybrid abutments). ....	93
<b>Table 12:</b> Results of the Tukey HSD Post Hoc test for fracture resistance mean values of PHAC and PHA (PHAC: Pressed hybrid abutment crowns, PHA: Pressed hybrid abutment).....	94

**Table 13:** Results of the Tukey HSD Post Hoc test for fracture resistance mean values of PHAC and CHA (PHAC: Pressed hybrid abutment crowns, CHA: Cemented hybrid abutments)..... 95

**Table 14:** Results of the Tukey HSD Post Hoc test for fracture resistance mean values of CHA and PHA (CHA: Cemented hybrid abutments, PHA: Pressed hybrid abutments). .... 96